AD	

Award Number: DAMD17-03-1-0101

TITLE: A Novel Approach to Monitoring Prostate Tumor

Oxygenation: Proton MRI of the Reporter Molecule

Hexamethyldisiloxane

PRINCIPAL INVESTIGATOR: Weina Cui, Ph.D.

CONTRACTING ORGANIZATION: University of Texas Southwestern Medical

Center

Dallas, Texas 75390-9105

REPORT DATE: March 2004

TYPE OF REPORT: Annual Summary

PREPARED FOR: U.S. Army Medical Research and Materiel Command

Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for Public Release;

Distribution Unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

REPORT DOCUMENTATION PAGE

Form Approved OMB No. 074-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503

1. AGENCY USE ONLY	2. REPORT DATE	3. REPORT TYPE AND DATES COVERED	
(Leave blank)	March 2004	Annual Summary (1 Mar 2003 - 28 Feb 200)4)

4. TITLE AND SUBTITLE
A Novel Approach to Monitoring Prostate Tumor
Oxygenation: Proton MRI of the Reporter Molecule
Hexamethyldisiloxane

5. FUNDING NUMBERS
DAMD17-03-1-0101

6. AUTHOR(S)

Weina Cui, Ph.D.

7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES)
University of Texas Southwestern Medical Center

Dallas, Texas 75390-9105

8. PERFORMING ORGANIZATION REPORT NUMBER

E-Mail: weina.cui@utsouthwestern.edu

9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES)

U.S. Army Medical Research and Materiel Command Fort Detrick, Maryland 21702-5012

10. SPONSORING / MONITORING AGENCY REPORT NUMBER

11. SUPPLEMENTARY NOTES

12a. DISTRIBUTION / AVAILABILITY STATEMENT

Approved for Public Release; Distribution Unlimited

12b. DISTRIBUTION CODE

13. ABSTRACT (Maximum 200 Words)

Growing evidence from experimental and clinical studies confirms that solid human tumors have foci of hypoxic cells, which have a profound influence on the therapeutic outcome of cancer chemotherapy and radiotherapy. A strong argument therefore exists for assessing the hypoxic fraction of tumors prior to patient treatment, and to tailor this treatment accordingly. It has been shown that there is linear relationship between R of hexamethyldisiloxane (HMDSO) and pO2, and the R1 of HMDSO is insensitive to various ions and minimally sensitive to temperature. The primary sequence for in vivo T1 measurment with water suppression has been established, and HMDSO is also detectable in prostate tumor. So, Hexamethyldisiloxane shows promise as a reporter molecule to measure tumor oxygenation by 1H MRS and potentially by MRI. This opens new opportunities for MR tumor oximetry, particularly since HMDSO is used widely in biomedical materials and as an ingredient in consumer products; and HMDSO is reported to have minimal toxicity.

14. SUBJECT TERMS			15. NUMBER OF PAGES
Hexamethyldisiloxane,	hypoxia, prostate tumo	r, proton MRI, oximetry	12
			16. PRICE CODE
17. SECURITY CLASSIFICATION OF REPORT	18. SECURITY CLASSIFICATION OF THIS PAGE	19. SECURITY CLASSIFICATION OF ABSTRACT	20. LIMITATION OF ABSTRACT
Unclassified	Unclassified	Unclassified	Unlimited

Table of Contents

Cover1
SF 2982
Table of Contents3
Introduction4
Body5
Key Research Accomplishments6
Reportable Outcomes9
Conclusions9
References9
Appendices11

Introduction:

Growing evidence from experimental and clinical studies confirms that solid human tumors have foci of hypoxic cells, which have a profound influence on the therapeutic outcome of cancer chemotherapy and radiotherapy: the level and severity of hypoxia can be a strong prognostic factor of disease progression and survival. A strong argument therefore exists for assessing the hypoxic fraction of tumors prior to patient treatment, and to tailor this treatment accordingly.

Baseline pO₂ and dynamic changes with respiratory intervention has been measured extensively in rat prostate tumors by ¹⁹F MRI(FREDOM, Fluorocarbon Relaxometry using Echo planar imaging for Dynamic Oxygen Mapping) based on the spin-lattice relaxation of Hexafluorobenzene(HFB), which provides spatially resolved maps of tumor oxygenation at depth and allows monitoring of dynamic changes at specific locations. However, for clinical application ¹⁹F NMR is not yet widely available. A proton MR analogue of HFB could facilitate immediate widespread oximetry. We have identified hexamethyldisiloxane(HMDSO) as a potential reporter. HMDSO has extensive symmetry and a single proton resonance well removed from water.

We had three specific aims for this project:

- **Task1.** Characterize the effects of external factors on T1(spin-lattice relaxation time) of HMDSO in vitro and get the calibration curves, Month 1-6
- **Task2.** Develop relevant MR pulse sequences with H₂O suppression for imaging. Establish experimental prostate tumor models and measure the retention time of HMDSO in tumors. Month 6-12
- **Task3.** Make pO_2 maps and measure prostate tumor oxygen dynamics with respect to growth rate and respiratory challenge, Month 12-24

Body:

Hexamethyldisiloxane(HMDSO) is used widely in biomedical materials and as an ingredient in consumer products, such as a thin polymeric coating on suture for cardiovascular surgery, or the thin layer onto the inner surface of plasma-modified small diameter tubing, etc. HMDSO is reported to have minimal toxicity. A previous test, by me, in vitro indicates that the spin-lattice relaxation time of HMDSO is sensitive to oxygen tension. HMDSO is highly hydrophobic, and therefore can be injected intratumorally at specific sites and will not diffuse in the tumor. Moreover, the boiling point of HMDSO is 99-100°C. In addition, HMDSO only has one ¹H signal and the chemical shift difference between HMDSO and H₂O is about 4.7ppm. Thus, we believe HMDSO may be appropriate for measuring tumor pO₂ by MRS and MRI of ¹H. In this proposal, I plan to develop proton MR techniques using this new reporter molecule, HMDSO, to assess the prostate tumor oxygenation in different prostate tumor sublines with several selected levels of histological differentiations.

$$H_3C$$
 H_3C
 Si
 O
 Si
 CH_3
 CH_3
 CH_3

For the first year, I had the statement of work as:

Task1. Characterize the effects of external factors on T1(spin-lattice relaxation time) of HMDSO in vitro and get the calibration curves, Month 1-6

Measure the influence of

- a. different temperatures on T1 of HMDSO
- b. different metal ions on T1 of HMDSO
- c. different B₀ on T1 of HMDSO
- d. different pO₂ on T1 of HMDSO at various temperature, metal ions and B₀.


Task2. Develop relevant MR pulse sequences with H₂O suppression for imaging. Establish experimental prostate tumor models and measure the retention time of HMDSO in tumors. Month 6-12

- a. Become proficient with state of the art NMR e.g. echo planar imaging
- b. Design and implement special pulse sequences to suppress the resonance of H_2O while meaning relaxation
- c. Become proficient with surgically creating tumor pedicles and implantation of prostate tumors
- d. Use three selected sublines of the Dunning prostate tumor in rats, allow to grow to 1 cm diameter and measure the retention time of HMDSO


Key research accomplishments:

- **Task1.** In order to characterize the effects of external factors on T1(spin-lattice relaxation time) of HMDSO in vitro and get the calibration curves, we measured the influence of: a). different ions on T1 of HMDSO, b). different temperatures on T1 of HMDSO, c) different B0 on T1 measurement, d) different pO2 on T1 measurement
 - 1. T1 of HMDSO with different oxygen tension was measured on Varian 200 MHz and 600 MHz.

200MHz:




600 MHz




From above data, we can conclude that there is linear relationship between oxygen tension and spin-lattice relaxation rate. The correlation equation at 200 MHz is: R1=0.15+0.00168x (X is oxygen tension (torr).

2. Prepare the solution of Na⁺, K⁺, Mg²⁺, Ca²⁺, the mixture of HMDSO+H2O, and the ion concentration is calculated according to the proportion of various ion in vivo. The

T1 of these solutions were measured on 200MHz with various temperature and oxygen tension. It was found that there is no significant influence of various ions on spin-lattice relaxation rate, especially in the condition of lower oxygen tension that tumors have. R1 of HMDSO is only minimally sensitive to temperature. The correlation equation at 37°C is R1=0.13+0.00175x (X is oxygen tension (torr).






Task2. Develop relevant MR pulse sequences with H₂O suppression for imaging. Establish experimental prostate tumor models and measure the retention time of HMDSO in tumors

a. I become proficient with surgically creating tumor pedicles and implantation of prostate tumors

For thigh tumor: a flap of depilated skin is raised from the thigh of young adult male Copenhagen rat and held in position. A 1cm incision is made. A piece of fresh tumor tissue (2x2x2mm³) is implanted under the skin and the cut closed with a wound clip.


For pedicle tumor: A flap of depilated skin is raised from the body of young adult male Copenhagen rat and held in position with a non-traumatic curved bull-dog clip. A 3cm incision is made through the skin using the curved edge of the clip as guide. Would clips are used to join the edge of the skin producing a tube resembling a suitcase handle. Two weeks later, the clips are removed and the distal end of the pedicle severed. A piece of fresh tumor tissue (2x2x2mm³) is implanted in the lumen and the cut closed with a wound clip.

b. I used selected sublines of the Dunning prostate tumor in rats to measure the retention time of HMDSO in tumor. 100 µ l HMDSO was injected intratumorally into Dunning prostate 3327 AT1 and H tumor, and PRESS pulse sequence with water suppression was used to acquire data. It was shown that HMDSO can stay at least 24hr within prostate tumor.



c. We design and implement special pulse sequence to suppress the resonance of H2O while meaning relaxation. This work supported by MR scientist Dr. Vikram Kodibagkar.

For T1 measurements by spectroscopy, a pulse burst saturation recovery (PBSR) sequence consisting of 20 nonselective 90° pulses for saturation of signal followed by a delay tau for magnetization recovery—and a single, frequency selective 90° pulse for signal detection was used. T1 values were measured using this sequence with the ARDVARC (Alternating Relaxation Delays with Variable Acquisitions for Reduction of Clearance effects) protocol. For imaging experiments a spin-echo EPI based PBSR pulse sequence is being designed for measuring T1 values. The sequence will consist of a) a train of 20 non-selective 90° pulses for saturation of signal followed by a delay tau for magnetization recovery, b) three optional CHESS pulses for frequency selective saturation of residual water and fat immediately followed by c) spin-echo EPI detection with a slice selective 90° pulse and a frequency selective 180° pulse.



Reportable outcomes:

One paper describing work supported by this grant will be presented as poster at the International Society of Magnetic Resonance in Medicine on May 2004.

Hexamethyldisiloxane (HMDSO), a novel reporter molecule for in-vivo oximetry using 1H MRI., Vikram D. Kodibagkar, Matthew E. Merritt, Weina Cui, Ralph P. Mason. ISMRM, May 15, 2004, Kyoto, Japan

Conclusions:

The above experiments showed that: 1) there is linear relationship between R1 of HMDSO and pO₂; 2) the R1 of HMDSO is insensitive to various ions and minimally sensitive to temperature; 3) the correlation equation at 37°C is R1=0.13+0.00175x (X is oxygen tension (torr),; 4) HMDSO is detectable in a tumor, and it remains present in tumor more than 24hrs; 5) the primary sequence for in vivo T1 measurement with water suppression has been established. So, Hexamethyldisiloxane shows promise as a reporter molecule to measure tumor oxygenation by ¹H MRS and potentially by MRI. This opens new opportunities for MR tumor oximetry, particularly since HMDSO is used widely in biomedical materials and as an ingredient in consumer products; and HMDSO is reported to have minimal toxicity.

References:

Hunjan S, Zhao D, Constantinescu A, Hahn EW, Antich PP, Mason RP., Tumor oximetry: demonstration of an enhanced dynamic mapping procedure using fluorine-19 echo planar

magnetic resonance imaging in the Dunning prostate R3327-AT1 rat tumor, Int J Radiat Oncol Biol Phys. 2001 Mar 15;49(4):1097-108

Le D, Mason RP, Hunjan S, Constantinescu A, Barker BR, Antich PP, Regional tumor oxygen dynamics: 19F PBSR EPI of hexafluorobenzene. Magn Reson Imaging. 1997;15(8):971-81.

Mason RP, Rodbumrung W, Antich PP., Hexafluorobenzene: a sensitive 19F NMR indicator of tumor oxygenation. NMR Biomed. 1996 May;9(3):125-34.

Appendix 1

Hexamethyldisiloxane (HMDSO), a novel reporter molecule for in-vivo oximetry using 1H MRI. Vikram D. Kodibagkar, Matthew E. Merritt, Weina Cui, Ralph P. Mason. Department of Radiology, UT Southwestern Medical Center, Dallas, TX

Synopsis

We present here, a proof of principle for the use of hexamethyldisiloxane (HMDSO) as a novel MRI based oximetry probe. The presence of water and fat resonances is the chief hurdle in imaging and relaxometry of a proton based oximetry probe. A spin-echo EPI based imaging protocol is presented with water and fat suppression for potential in vivo relaxometry applications. Results of phantom studies are reported.


Introduction


There is increasing evidence for the importance of tissue oxygenation in development, progression, and response to cancer therapy. Oxygen is required for efficient function by most tissues and hypoxia leads to rapid cellular dysfunction and damage. In addition, hypoxic tumor cells are refractory to radiotherapy. Thus, the opportunity to measure tissue oxygen tension non-invasively may be significant in understanding mechanisms of tissue function and in clinical prognosis. The linear dependence of R_1 of fluorocarbon F resonances on pO_2 is well known and has been studied extensively. We have studied the potential of HMDSO as a H based pO_2 reporter molecule. In a related study (submitted to ISMRM 2004) data is presented for the linear dependence of R_1 of HMDSO on pO_2 (R_1 = 0.13+ 0.00175*pO2[torr] at 37 °C) and the relative insensitivity of R_1 to temperature fluctuations and presence of physiologically relevant metal ions. Here we demonstrate the implementation of an imaging protocol (suitable for future in vivo studies) on phantoms.


Materials and Methods


The data was acquired on a Varian 4.7T scanner. A spin-echo EPI based pulse sequence was used for imaging and measuring T₁ values. The sequence consisted of a) a train of 20 non-selective 90° pulses for saturation of signal followed by a delay *tau* for magnetization recovery, b) 3 CHESS¹ pulses for frequency selective saturation of water and fat immediately followed by c) spin-echo EPI detection with a slice selective 90° pulse and a frequency selective 180° pulse. T₁ values were measured using this sequence with the ARDVARC (Alternating Relaxation Delays with Variable Acquisitions for Reduction of Clearance effects) protocol . For comparison, reference images were obtained using a spin echo sequence. T₁ maps were made using the Varian Image Browser software.

Results and Discussion









a) T_l weighted spin-echo image of phantom with smaller tubes containing mineral oil (F) and HMDSO (H) inside a tube containing water (W) and b) proton density weighted EPI image of the same phantom with fat and water suppression. T_l maps of a phantom containing HMDSO saturated with gases with different concentrations of oxygen obtained by c) spin-echo sequence and d) the spin-echo EPI sequence.

HMDSO has a single proton resonance with a chemical shift of –5ppm relative to water. Since our application is oriented towards T1 relaxometry, our approach to suppression involves frequency selective pulses. The EPI based sequence with frequency selective excitation and suppression was demonstrated to effectively suppress signals from mineral oil (to simulate fat) and water. The choice of a long echo time (>100ms) enabled suppression of any residual fat signal, resulting from its proximity to the HMDSO resonance. T1 data was obtained within 2 and half minutes. This approach is tailored towards future in vivo applications and the short acquisition time can allow us to monitor dynamic response to oxygen challenge. Minimal toxicity and wide availability add to the promise of HMDSO as a pO2 reporter molecule.

Acknowledgements

This work was supported by NCI Pre-ICMIC P20 CA086354 and DOD DAMD17-03-1-0101

References

- (1) Haase, A.; Frahm, J.; Hanicke, W.; Matthaei, D. Phys Med Biol 1985, 30, 341-4.
- (2) Hunjan, S.; Zhao, D.; Constantinescu, A.; Hahn, E. W.; Antich, P. P.; Mason, R. P. Int. J. Radiat. Oncol. Biol. Phys. 2001, 49, 1097-1108.